



**Madison County**  
**Department of Weights and Measures**  
 138 N. Court St. | Building 5  
 Wampsville, NY 13163  
 315.366.2860

Test # 21104

White - Department  
 Yellow - Owner  
 Pink - Municipal Office

**Large Capacity Scale Report**

Jurisdiction Madison County		Date 7/25/24	
Establishment Name NY Hay Sales		Phone 315-794-3909	Business Type Agriculture
Address 7149 Hoose Rd		Scale Cap x d 200,000 x 20lb	CLC/Sect Cap 70,000lbs
City, State, Zip Earlville, NY 13332		WM Lic. (Y/N/NA)	
Indicator Type Digital	Mfr. B-Tek	Model T403S	S/N 202450312
Platform Type Load Cell	Mfr. B-Tek	Model BT-7011-200-FESD-SL	S/N TS-08218790
Deck Type/Length steel/70'		# Sections 4	Approaches Gravel/Concrete
			Overall Condition Good

1 Test Type	2 Strain Load Reading	3 Sub Reading	4 Error @ Sub Load From Col 9	5 Sub Load Col(3 - 4)	6 Known Weight Added	7 Expected Reading Col (2+5+6)	8 Indicator Reading	9 Error Col (8 - 7)	Section Tests
ILZ					3000	3000	3000	0	Test Load 24,000
					3000	6000	6000	0	Reading / Error
					3000	9000	9000	0	1. 24,000 / 0
					3000	12000	12,000	0	2. 24,000 / 0
					3000	15,000	15,000	0	3. 23,920 / -20
					3000	18,000	17,980	-20	4. 23,920 / -20
					3000	21,000	21,000	0	5. H-L / 20
					3000	24,000	24,000	0	6. /
					0	0	0	0	7. /
SUB#1 w/ Truck	26920	0	26920	24,000	50,920	50,920	0	0	Test Load
					0	0	0	0	Reading / Error
SUB#2 Tractor	50,820	0	50,820	50,920	101,740	101,740	0	0	1. /
					50,920	50,920	50,920	0	2. /
DL					0	0	0	0	3. /
									4. /
									5. /
									6. /
									7. /

Sensitivity Load @ 0 \_\_\_\_\_ Sensitivity Load @ Max Test Load \_\_\_\_\_ Weights furnished by Freeman Scale last certified 9/21

Test Type: **ILZ** - Increasing Load @ Zero    **SUB#** - Substitution    **STR** - Strain Load    **DL** - Decreasing Load

<input checked="" type="checkbox"/> Normal Test	<b>Tolerance Applied</b>	<b>Official Action</b>
<input type="checkbox"/> Retest (following rejection/repair)	<input checked="" type="checkbox"/> Maintenance	<input checked="" type="checkbox"/> Approved/Sealed
<input type="checkbox"/> Retest (following adjustment @ site)	<input type="checkbox"/> Acceptance	<input type="checkbox"/> Rejected/Ordered Repaired

Remarks Scale tested within specified tolerances  
L-4, CA1-7

Acknowledged 	Inspected by <u>Mark Platte</u>	Inspector Number <u>195</u>
------------------	------------------------------------	--------------------------------